



# Korea Environment & Merchandise Testing Institute

459-28 GASAN-DONG, GEUMCHEON-GU, SEOUL, KOREA TEL : 02)856-5615

FAX : 02)856-5618, 854-6667 <http://www.kemti.org>

## TEST REPORT

No. : 10949

Client : CHANG YOUNG CHEMICAL

Address : 536-70, Gumkang-Ri, Gumkang-Myun, Anseong-Shi,  
Kyeonggi-do, Korea

Date Out : JUN. 01. 2004

Date In : MAY. 24. 2004

Test Date : JUN. 01. 2004

Sample Description : PVC Compound(For INFUSION SET TUBE)

Test Results :

Items	Units	Criteria	Results
Appearance	-	Not be turbidity, coloration or particles.	Conforms
pH(at 25°C)	-	Difference in pH : Max. 1.5	0.3
Reducing Agents	ml	Difference in titres : Max. 2.0	0.5
Total Extractables	mg	Difference in extractables : Max. 1.0	0.2
Heavy Metals (as Pb)	-	Shall not be darker than color of the Standard Preparation.	Conforms
Test Method	KFDA NOTICE NO.2002-55, Blood donor set and Disposable Intravascular administration set		

Signed by

Soo-Boo Chun, Manager  
Toxicology Lab.

OUR REPORTS APPLY ONLY TO THE STANDARDS OR PROCEDURES IDENTIFIED AND TO THE SAMPLE(S) TESTED UNLESS OTHERWISE SPECIFIED. THE TEST RESULTS ARE NOT INDICATIVE OF REPRESENTATIVE OF THE QUALITIES OF THE LOT FROM WHICH THE SAMPLE WAS TAKEN OR OF APPARENTLY IDENTICAL OR SIMILAR PRODUCTS.